



MS4000-S | VOC Gas Analyzer | 1 - 3000 PPM

The **MS4000-S** is a **VOC and hydrocarbon gas concentration analyzer**. It is designed to monitor VOC gases in industrial processes and applications as well as in environmental monitoring applications.

The instrument senses gases or volatiles providing a measurement system with very low maintenance requirements.

APPLICATIONS

- Monitoring of Total VOC
- Process VOC Monitoring
- Environmental Monitoring
- Detection of VOCs in wastewater systems
- Odor Control
- Sewage Network Monitoring
- Natural Gas Extraction

FEATURES AND BENEFITS

- Continuous sampling and measurement
- Built-in humidity compensation
- Relay control of users' alarms
- Integrated data logging
- Versatile communication protocols
- Powerful diagnostics



- ✓ Continuous monitoring
- ✓ Long sensor lifetime: 5+ years*
- ✓ Low operating and maintenance costs
- ✓ Ideal for the industrial sector
- ✓ Low maintenance, no sensor cleaning
- ✓ Field proven technology

TECHNICAL SPECIFICATION

PARAMETER	OPERATIONAL REQUIREMENTS		NOTES
	MINIMUM	MAXIMUM	
DETECTION RANGE	1 ppm	3000 ppm	Total VOCs
4-20 mA RANGE (DEFAULT)	0 ppm	3000 ppm	Configurable on commissioning
REPEATABILITY	-5%	+5%	Measured against a 100 ppm toluene standard at standard temperature and pressure; accuracy can be further improved at concentrations of interest, using bespoke calibration available on request.
ACCURACY	-15%	+15%	
MEASUREMENT FREQUENCY	Continuous		
AMBIENT TEMPERATURE	0 °C	50 °C	Climate control might be needed depending on process and ambient conditions
RELATIVE HUMIDITY OF SAMPLE		90%	Non-condensing
ANALOG INPUT	4 mA	20 mA	Scalable to range required, max load 750 Ω, loop powered or instrument powered
ANALOG OUTPUT	4 mA	20 mA	
ANALOG OUTPUT ISOLATION	5 kVAC, 400 VDC		
RELAY CURRENT		5 A	Max. 250 VAC or 30 VDC; 3x: Alarm 1, 2 and Fault Relays with NO and NC contacts
INTERFACE	Touchscreen		With advanced reporting capabilities
COMMUNICATION PROTOCOLS	Modbus (Optional)		Profibus / HART / Foundation Fieldbus on request
DATA STORAGE	μSD Card / Windows PC		
INSTRUMENT CASE	IP65		Powder coated stainless steel; Power Supply Module IP68
INSTRUMENT WEIGHT	15 kg		
SYSTEM DIMENSIONS	45 x 28 x 20 cm		Power Supply module mounted externally next to instrument.
SUPPLY VOLTAGE	90 VAC	240 VAC	50 Hz or 60 Hz
POWER CONSUMPTION		25 W	

*Under standard operating conditions

TYPICAL TARGET SUBSTANCES

Petroleum Gas, Methane, Iso Butane, Kerosene, Ethanol, Diesel, Iso-butylene, Toluene, Acetone, Xylene, Heptane, Benzene, Methanol, Mercaptan



CONSUMABLES

6 Months or as required: Filters and restrictors

SERVICE

Service is necessary every 6 months to keep the instrument operating in optimal conditions. Air Pump replaced every 18-36 months as required. Service should be carried out by a trained technician who has undergone the standard Multisensor Systems training course.



**HEAD OFFICE
(UNITED KINGDOM)**

Multisensor Systems Ltd.

Alexandra Court
Carrs Road
Cheadle
SK8 2JY

United Kingdom

T: +44 (0)161 491 5600

E: info@multisensorsystems.com

USA - WEST COAST

RealTech Controls

Los Angeles, California

W: www.realtechcontrols.com

E: support@realtechcontrols.com

T: (818) 606-0562

USA - SOUTH EAST

Volition Controls Corp.

715 Tillman Place, Suite 104

Plant City, FL 33566

W: www.volitioncc.com

E: sales@volitioncc.com

T: 813-714-7006

CANADA

WJF Instrumentation Ltd.

#5 3610 – 29th Street N.E.

Calgary, Alberta, T1Y 5Z7

W: www.wjf.ca

E: info@wjf.ca

T: 877 291 5572



Multisensor Systems Limited reserves the right to revise any specifications and data contained within this document without notice.

Multisensor Systems is a developer and supplier of Water and Gas Analysers specialising in oil in water and hydrocarbon analysers, oil in water detectors, VOC monitors and THM analysers based in the United Kingdom.

The contents of this publication are provided to you "as is" without warranty of any kind, and are subject to change without notice. Multisensor Systems does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.

Multisensor Systems Ltd., Alexandra Court, Carrs Road, Cheadle, SK8 2JY, United Kingdom

©2010-Present, Multisensor Systems Limited

CHANGELOG

MSS DOCUMENT CHANGE RECORD

DOCUMENT REF 1-000435

DATE	VERSION	CHANGED BY	CHECKED BY	ECN REF
19/09/2025	1.0	GO	LR	0925-03
23/09/2025	1.1	GO	LR	0925-04
25/09/2025	1.2	GO	LR	0925-07